

358 Mass Flow Meter & 1358 Mass Flow Controller Pin-outs

Pin	Description
1	Valve Test Point
2	Flow Signal Output
3	Valve Close (low)
4	Valve Open (low)
5	Power Common
6	-15 VDC
7	+15 VDC
8	Set Point Input
9	No Connection
10	Optional Input
11	Signal Common
12	Signal Common
13-14	No Connection
15	Chassis Ground.