


907 Analog Convection Transducer Pin-outs



Pin	Description
1	Set Point Relay - Normally Open Contact
2	Set Point Relay - Normally Closed Contact
3	Power Supply Input (+)
4	Power Supply Input (-)
5	Analog Output Voltage (+)
6	Set Point Relay - Common
7	Set Point Relay - Disable
8	Analog Output Voltage (-)
9	Set point relay output voltage

Dual Set Point Version
15-pin High Density D-sub

Pin	Description
1	Set point relay 1 output voltage
2	Set point relay 2 output voltage
3	Power supply input (+)
4	Power supply input (-)
5	Analog output voltage (+)
6	Analog output voltage (-)
7	Not used
8	Not used
9	Not used
10	Set point relay 1 - normally open contact
11	Set point relay 2 - normally open contact

12	Set point relay 2 - common
13	Set point relay 2 - normally closed contact
14	Set point relay 1 - normally closed contact
15	Set point relay 1 - common